

ANALYTICAL REPORT

Montana Environmental Laboratory LLC
1170 N. Meridian Rd., P.O. Box 8900, Kalispell, MT 59904-1900
Phone: 406-755-2131 Fax: 406-257-5359 www.melab.us

Randy Sundberg
Many Lakes Vacation Village HOA
P.O. Box 2032
Bigfork, MT 59911

PWS ID:
Project:

Client Sample ID: Plumers Lake **Lab ID:** 2206823-01
Matrix: ENVIRO WATER **Collected:** 07/01/2022 8:05 **Received:** 07/01/2022 13:30

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	1.08	mg/L	0.1		SM4500NOrg		07/12/2022 13:50	NB
Phosphorus, Total low level	0.019	mg/L	0.001		E365.1		07/06/2022 13:18	NB
Turbidity	1.25	NTU	0.05		E180.1		07/01/2022 17:00	BLW

Client Sample ID: Kathy Lake SW **Lab ID:** 2206823-02
Matrix: ENVIRO WATER **Collected:** 07/01/2022 8:19 **Received:** 07/01/2022 13:30


<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	0.72	mg/L	0.1		SM4500NOrg		07/12/2022 13:50	NB
Phosphorus, Total low level	0.013	mg/L	0.001		E365.1		07/06/2022 13:18	NB
Turbidity	0.52	NTU	0.05		E180.1		07/01/2022 17:00	BLW

Client Sample ID: Kathy Lake NE **Lab ID:** 2206823-03
Matrix: ENVIRO WATER **Collected:** 07/01/2022 8:25 **Received:** 07/01/2022 13:30

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	0.71	mg/L	0.1		SM4500NOrg		07/12/2022 13:50	NB
Phosphorus, Total low level	0.014	mg/L	0.001		E365.1		07/06/2022 13:18	NB
Turbidity	0.45	NTU	0.05		E180.1		07/01/2022 17:00	BLW

Client Sample ID: Amy Lake S **Lab ID:** 2206823-04
Matrix: ENVIRO WATER **Collected:** 07/01/2022 8:29 **Received:** 07/01/2022 13:30

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	0.77	mg/L	0.1		SM4500NOrg		07/12/2022 13:50	NB
Phosphorus, Total low level	0.020	mg/L	0.001		E365.1		07/06/2022 13:18	NB
Turbidity	0.47	NTU	0.05		E180.1		07/01/2022 17:00	BLW



ANALYTICAL REPORT

Montana Environmental Laboratory LLC
1170 N. Meridian Rd., P.O. Box 8900, Kalispell, MT 59904-1900
Phone: 406-755-2131 Fax: 406-257-5359 www.melab.us

Randy Sundberg
Many Lakes Vacation Village HOA
P.O. Box 2032
Bigfork, MT 59911

PWS ID:
Project:

Client Sample ID: Ann Lake W **Lab ID:** 2206823-05
Matrix: ENVIRO WATER **Collected:** 07/01/2022 8:36 **Received:** 07/01/2022 13:30

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	0.70	mg/L	0.1		SM4500NOrg		07/12/2022 13:50	NB
Phosphorus, Total low level	0.015	mg/L	0.001		E365.1		07/06/2022 13:18	NB
Turbidity	0.38	NTU	0.05		E180.1		07/01/2022 17:00	BLW

Client Sample ID: John Lake W **Lab ID:** 2206823-06
Matrix: ENVIRO WATER **Collected:** 07/01/2022 8:45 **Received:** 07/01/2022 13:30


<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	0.44	mg/L	0.1		SM4500NOrg		07/12/2022 13:50	NB
Phosphorus, Total low level	0.009	mg/L	0.001		E365.1		07/13/2022 13:49	NB
Turbidity	0.38	NTU	0.05		E180.1		07/01/2022 17:00	BLW

Client Sample ID: John Lake E **Lab ID:** 2206823-07
Matrix: ENVIRO WATER **Collected:** 07/01/2022 8:52 **Received:** 07/01/2022 13:30

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	0.50	mg/L	0.1		SM4500NOrg		07/12/2022 13:50	NB
Phosphorus, Total low level	0.009	mg/L	0.001		E365.1		07/13/2022 13:49	NB
Turbidity	0.37	NTU	0.05		E180.1		07/01/2022 17:00	BLW

Client Sample ID: Cree Lake W **Lab ID:** 2206823-08
Matrix: ENVIRO WATER **Collected:** 07/01/2022 8:57 **Received:** 07/01/2022 13:30

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	0.62	mg/L	0.1		SM4500NOrg		07/12/2022 13:50	NB
Phosphorus, Total low level	0.007	mg/L	0.001		E365.1		07/13/2022 13:49	NB
Turbidity	0.37	NTU	0.05		E180.1		07/01/2022 17:00	BLW



ANALYTICAL REPORT

Montana Environmental Laboratory LLC

1170 N. Meridian Rd., P.O. Box 8900, Kalispell, MT 59904-1900

Phone: 406-755-2131 Fax: 406-257-5359 www.melab.us

Randy Sundberg
Many Lakes Vacation Village HOA
P.O. Box 2032
Bigfork, MT 59911

PWS ID:
Project:

Client Sample ID: Swimming Lake E
Matrix: ENVIRO WATER

Collected: 07/01/2022 9:00

Lab ID: 2206823-09
Received: 07/01/2022 13:30

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	0.65	mg/L	0.1		SM4500NOrg		07/12/2022 13:50	NB
Phosphorus, Total low level	0.007	mg/L	0.001		E365.1		07/13/2022 13:49	NB
Turbidity	0.81	NTU	0.05		E180.1		07/01/2022 17:00	BLW

Client Sample ID: Swimming Lake S
Matrix: ENVIRO WATER

Collected: 07/01/2022 9:06

Lab ID: 2206823-10
Received: 07/01/2022 13:30

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	0.64	mg/L	0.1		SM4500NOrg		07/12/2022 13:50	NB
Phosphorus, Total low level	0.008	mg/L	0.001		E365.1		07/13/2022 13:49	NB
Turbidity	0.51	NTU	0.05		E180.1		07/01/2022 17:00	BLW

Client Sample ID: Gilbert Lake W
Matrix: ENVIRO WATER

Collected: 07/01/2022 9:10

Lab ID: 2206823-11
Received: 07/01/2022 13:30


<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	0.67	mg/L	0.1		SM4500NOrg		07/12/2022 13:50	NB
Phosphorus, Total low level	0.008	mg/L	0.001		E365.1		07/13/2022 13:49	NB
Turbidity	0.38	NTU	0.05		E180.1		07/01/2022 17:00	BLW

Client Sample ID: Jake Lake
Matrix: ENVIRO WATER

Collected: 07/01/2022 9:15

Lab ID: 2206823-12
Received: 07/01/2022 13:30

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	0.46	mg/L	0.1		SM4500NOrg		07/12/2022 13:50	NB
Phosphorus, Total low level	0.007	mg/L	0.001		E365.1		07/13/2022 13:49	NB
Turbidity	0.39	NTU	0.05		E180.1		07/01/2022 17:00	BLW



ANALYTICAL REPORT

Montana Environmental Laboratory LLC
1170 N. Meridian Rd., P.O. Box 8900, Kalispell, MT 59904-1900
Phone: 406-755-2131 Fax: 406-257-5359 www.melab.us

Randy Sundberg
Many Lakes Vacation Village HOA
P.O. Box 2032
Bigfork, MT 59911

PWS ID:
Project:

Client Sample ID: Gilbert Lake E **Lab ID:** 2206823-13
Matrix: ENVIRO WATER **Collected:** 07/01/2022 9:24 **Received:** 07/01/2022 13:30

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	0.64	mg/L	0.1		SM4500NOrg		07/12/2022 13:50	NB
Phosphorus, Total low level	0.008	mg/L	0.001		E365.1		07/13/2022 13:49	NB
Turbidity	0.36	NTU	0.05		E180.1		07/01/2022 17:00	BLW

Client Sample ID: Harbin Lake E **Lab ID:** 2206823-14
Matrix: ENVIRO WATER **Collected:** 07/01/2022 9:37 **Received:** 07/01/2022 13:30


<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	0.27	mg/L	0.1		SM4500NOrg		07/12/2022 13:50	NB
Phosphorus, Total low level	0.006	mg/L	0.001		E365.1		07/13/2022 13:49	NB
Turbidity	0.30	NTU	0.05		E180.1		07/01/2022 17:00	BLW

Client Sample ID: Lake N of Kid W **Lab ID:** 2206823-15
Matrix: ENVIRO WATER **Collected:** 07/01/2022 9:50 **Received:** 07/01/2022 13:30

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	0.71	mg/L	0.1		SM4500NOrg		07/12/2022 13:50	NB
Phosphorus, Total low level	0.027	mg/L	0.001		E365.1		07/13/2022 13:49	NB
Turbidity	0.54	NTU	0.05		E180.1		07/01/2022 17:00	BLW

Client Sample ID: Kid Lake **Lab ID:** 2206823-16
Matrix: ENVIRO WATER **Collected:** 07/01/2022 9:55 **Received:** 07/01/2022 13:30

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	0.55	mg/L	0.1		SM4500NOrg		07/12/2022 13:50	NB
Phosphorus, Total low level	0.009	mg/L	0.001		E365.1		07/13/2022 13:49	NB
Turbidity	0.30	NTU	0.05		E180.1		07/01/2022 17:00	BLW



ANALYTICAL REPORT

Montana Environmental Laboratory LLC
1170 N. Meridian Rd., P.O. Box 8900, Kalispell, MT 59904-1900
Phone: 406-755-2131 Fax: 406-257-5359 www.melab.us

Randy Sundberg
Many Lakes Vacation Village HOA
P.O. Box 2032
Bigfork, MT 59911

PWS ID:
Project:

Client Sample ID: Ann Lake E **Lab ID:** 2206823-17
Matrix: ENVIRO WATER **Collected:** 07/01/2022 10:04 **Received:** 07/01/2022 13:30

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	0.74	mg/L	0.1		SM4500NOrg		07/12/2022 13:50	NB
Phosphorus, Total low level	0.017	mg/L	0.001		E365.1		07/13/2022 13:49	NB
Turbidity	0.32	NTU	0.05		E180.1		07/01/2022 17:00	BLW

Client Sample ID: Douglas Lake NE **Lab ID:** 2206823-18
Matrix: ENVIRO WATER **Collected:** 07/01/2022 10:15 **Received:** 07/01/2022 13:30


<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	1.20	mg/L	0.1		SM4500NOrg		07/12/2022 13:50	NB
Phosphorus, Total low level	0.016	mg/L	0.001		E365.1		07/13/2022 13:49	NB
Turbidity	0.88	NTU	0.05		E180.1		07/01/2022 17:00	BLW

Client Sample ID: Douglas Lake W **Lab ID:** 2206823-19
Matrix: ENVIRO WATER **Collected:** 07/01/2022 10:20 **Received:** 07/01/2022 13:30

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	1.35	mg/L	0.1		SM4500NOrg		07/12/2022 13:50	NB
Phosphorus, Total low level	0.021	mg/L	0.001		E365.1		07/13/2022 13:49	NB
Turbidity	1.14	NTU	0.05		E180.1		07/01/2022 17:00	BLW

Client Sample ID: Peter Lake E **Lab ID:** 2206823-20
Matrix: ENVIRO WATER **Collected:** 07/01/2022 10:22 **Received:** 07/01/2022 13:30

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	1.10	mg/L	0.1		SM4500NOrg		07/12/2022 13:50	NB
Phosphorus, Total low level	0.012	mg/L	0.001		E365.1		07/13/2022 13:49	NB
Turbidity	1.05	NTU	0.05		E180.1		07/01/2022 17:00	BLW



ANALYTICAL REPORT

Montana Environmental Laboratory LLC
 1170 N. Meridian Rd., P.O. Box 8900, Kalispell, MT 59904-1900
 Phone: 406-755-2131 Fax: 406-257-5359 www.melab.us

Randy Sundberg
 Many Lakes Vacation Village HOA
 P.O. Box 2032
 Bigfork, MT 59911

PWS ID:
 Project:

Client Sample ID: Plummers Lake E
Matrix: ENVIRO WATER

Collected: 07/01/2022 10:30

Lab ID: 2206823-21
Received: 07/01/2022 13:30

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	0.99	mg/L	0.1		SM4500NOrg		07/19/2022 14:08	NB
Phosphorus, Total low level	0.013	mg/L	0.001		E365.1		07/13/2022 13:49	NB
Turbidity	1.02	NTU	0.05		E180.1		07/01/2022 17:00	BLW

Client Sample ID: Dan Lake NE
Matrix: ENVIRO WATER

Collected: 07/01/2022 10:35

Lab ID: 2206823-22
Received: 07/01/2022 13:30

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	3.27	mg/L	0.1		SM4500NOrg		07/19/2022 14:08	NB
Phosphorus, Total	0.47	mg/L	0.01		E365.1		07/13/2022 16:08	NB
Turbidity	2.60	NTU	0.05		E180.1		07/01/2022 17:00	BLW

Client Sample ID: Dan Lake SW
Matrix: ENVIRO WATER

Collected: 07/01/2022 10:40

Lab ID: 2206823-23
Received: 07/01/2022 13:30


<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	2.84	mg/L	0.1		SM4500NOrg		07/19/2022 14:08	NB
Phosphorus, Total	0.39	mg/L	0.01		E365.1		07/13/2022 16:08	NB
Turbidity	1.60	NTU	0.05		E180.1		07/01/2022 17:00	BLW

Client Sample ID: Hunnicutt Lake NE
Matrix: ENVIRO WATER

Collected: 07/01/2022 10:45

Lab ID: 2206823-24
Received: 07/01/2022 13:30

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	3.51	mg/L	0.1		SM4500NOrg		07/19/2022 14:08	NB
Phosphorus, Total	0.50	mg/L	0.01		E365.1		07/13/2022 16:08	NB
Turbidity	2.10	NTU	0.05		E180.1		07/01/2022 17:00	BLW



ANALYTICAL REPORT

Montana Environmental Laboratory LLC

1170 N. Meridian Rd., P.O. Box 8900, Kalispell, MT 59904-1900
Phone: 406-755-2131 Fax: 406-257-5359 www.melab.us

Randy Sundberg
Many Lakes Vacation Village HOA
P.O. Box 2032
Bigfork, MT 59911

PWS ID:
Project:

Client Sample ID: Peter Lake W **Lab ID:** 2206823-25
Matrix: ENVIRO WATER **Collected:** 07/01/2022 10:53 **Received:** 07/01/2022 13:30

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	1.01	mg/L	0.1		SM4500NOrg		07/19/2022 14:08	NB
Phosphorus, Total low level	0.013	mg/L	0.001		E365.1		07/13/2022 13:49	NB
Turbidity	1.00	NTU	0.05		E180.1		07/01/2022 17:00	BLW

Client Sample ID: Tamarack Lake N **Lab ID:** 2206823-26
Matrix: ENVIRO WATER **Collected:** 07/01/2022 10:58 **Received:** 07/01/2022 13:30


<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	0.61	mg/L	0.1		SM4500NOrg		07/19/2022 14:08	NB
Phosphorus, Total low level	0.007	mg/L	0.001		E365.1		07/13/2022 13:49	NB
Turbidity	0.82	NTU	0.05		E180.1		07/01/2022 17:00	BLW

Client Sample ID: Tamarack Lake SW **Lab ID:** 2206823-27
Matrix: ENVIRO WATER **Collected:** 07/01/2022 11:05 **Received:** 07/01/2022 13:30

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	0.60	mg/L	0.1		SM4500NOrg		07/19/2022 14:08	NB
Phosphorus, Total low level	0.007	mg/L	0.001		E365.1		07/13/2022 13:49	NB
Turbidity	0.91	NTU	0.05		E180.1		07/01/2022 17:00	BLW

Client Sample ID: Bass Lake SE **Lab ID:** 2206823-28
Matrix: ENVIRO WATER **Collected:** 07/01/2022 11:14 **Received:** 07/01/2022 13:30

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	2.18	mg/L	0.1		SM4500NOrg		07/19/2022 14:08	NB
Phosphorus, Total low level	0.063	mg/L	0.001		E365.1		07/13/2022 13:49	NB
Turbidity	9.20	NTU	0.05		E180.1		07/01/2022 17:00	BLW



ANALYTICAL REPORT

Montana Environmental Laboratory LLC

1170 N. Meridian Rd., P.O. Box 8900, Kalispell, MT 59904-1900
Phone: 406-755-2131 Fax: 406-257-5359 www.melab.us

Randy Sundberg
Many Lakes Vacation Village HOA
P.O. Box 2032
Bigfork, MT 59911

PWS ID:
Project:

Client Sample ID: Double Lake N **Lab ID:** 2206823-29
Matrix: ENVIRO WATER **Collected:** 07/01/2022 11:20 **Received:** 07/01/2022 13:30

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	0.89	mg/L	0.1		SM4500NOrg		07/19/2022 14:08	NB
Phosphorus, Total low level	0.009	mg/L	0.001		E365.1		07/13/2022 13:49	NB
Turbidity	0.60	NTU	0.05		E180.1		07/01/2022 17:00	BLW

Client Sample ID: E Bass Lake S **Lab ID:** 2206823-30
Matrix: ENVIRO WATER **Collected:** 07/01/2022 11:30 **Received:** 07/01/2022 13:30


<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	0.80	mg/L	0.1		SM4500NOrg		07/19/2022 14:08	NB
Phosphorus, Total low level	0.010	mg/L	0.001		E365.1		07/20/2022 12:33	NB
Turbidity	0.72	NTU	0.05		E180.1		07/01/2022 17:00	BLW

Client Sample ID: Double Lake E **Lab ID:** 2206823-31
Matrix: ENVIRO WATER **Collected:** 07/01/2022 11:36 **Received:** 07/01/2022 13:30

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	0.85	mg/L	0.1		SM4500NOrg		07/19/2022 14:08	NB
Phosphorus, Total low level	0.012	mg/L	0.001		E365.1		07/20/2022 12:33	NB
Turbidity	0.54	NTU	0.05		E180.1		07/01/2022 17:00	BLW

Client Sample ID: Lake SE of Double Lake **Lab ID:** 2206823-32
Matrix: ENVIRO WATER **Collected:** 07/01/2022 11:40 **Received:** 07/01/2022 13:30

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	1.11	mg/L	0.1		SM4500NOrg		07/19/2022 14:08	NB
Phosphorus, Total low level	0.018	mg/L	0.001		E365.1		07/20/2022 12:33	NB
Turbidity	0.78	NTU	0.05		E180.1		07/01/2022 17:00	BLW



ANALYTICAL REPORT

Montana Environmental Laboratory LLC
1170 N. Meridian Rd., P.O. Box 8900, Kalispell, MT 59904-1900
Phone: 406-755-2131 Fax: 406-257-5359 www.melab.us

Randy Sundberg
Many Lakes Vacation Village HOA
P.O. Box 2032
Bigfork, MT 59911

PWS ID:
Project:

Client Sample ID: Double Lake SW **Lab ID:** 2206823-33
Matrix: ENVIRO WATER **Collected:** 07/01/2022 11:49 **Received:** 07/01/2022 13:30

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	0.96	mg/L	0.1		SM4500NOrg		07/19/2022 14:08	NB
Phosphorus, Total low level	0.013	mg/L	0.001		E365.1		07/20/2022 12:33	NB
Turbidity	1.00	NTU	0.05		E180.1		07/01/2022 17:00	BLW

Client Sample ID: Vicki Lake N **Lab ID:** 2206823-34
Matrix: ENVIRO WATER **Collected:** 07/01/2022 11:54 **Received:** 07/01/2022 13:30

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	0.64	mg/L	0.1		SM4500NOrg		07/19/2022 14:08	NB
Phosphorus, Total low level	0.009	mg/L	0.001		E365.1		07/20/2022 12:33	NB
Turbidity	0.60	NTU	0.05		E180.1		07/01/2022 17:00	BLW

Client Sample ID: Vicki Lake SW **Lab ID:** 2206823-35
Matrix: ENVIRO WATER **Collected:** 07/01/2022 12:00 **Received:** 07/01/2022 13:30

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>RL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Nitrogen, Total Persulfate	0.70	mg/L	0.1		SM4500NOrg		07/19/2022 14:08	NB
Phosphorus, Total low level	0.011	mg/L	0.001		E365.1		07/20/2022 12:33	NB
Turbidity	0.74	NTU	0.05		E180.1		07/01/2022 17:00	BLW

QUALITY CONTROL REPORT

Montana Environmental Laboratory LLC

Test Parameters

Order#: M2206823

BLANK ENVIRO WATER	LAB-ID #	Sample Concentr.	Spike Concentr.	QC Test Result	Pct (%) Recovery	RPD
Nitrogen, Total Persulfate-mg/L	0018543-02			0		
Nitrogen, Total Persulfate-mg/L	0018565-02			0		
Phosphorus, Total-mg/L	0018547-02			0		
Phosphorus, Total-mg/L	0018547-02			0		
Phosphorus, Total low level-mg/L	0018531-02			0		
Phosphorus, Total low level-mg/L	0018548-02			0		
Phosphorus, Total low level-mg/L	0018548-02			0		
Phosphorus, Total low level-mg/L	0018566-02			0		
CONTROL ENVIRO WATER	LAB-ID #	Sample Concentr.	Spike Concentr.	QC Test Result	Pct (%) Recovery	RPD
Nitrogen, Total Persulfate-mg/L	0018543-03		1	0.99	99.%	
Nitrogen, Total Persulfate-mg/L	0018565-03		1	0.96	96.%	
Phosphorus, Total-mg/L	0018547-03		2	2.02	101.%	
Phosphorus, Total-mg/L	0018547-03		2	2.02	101.%	
Phosphorus, Total low level-mg/L	0018531-03		0.025	0.025	100.%	
Phosphorus, Total low level-mg/L	0018548-03		0.025	0.025	100.%	
Phosphorus, Total low level-mg/L	0018548-03		0.025	0.025	100.%	
Phosphorus, Total low level-mg/L	0018566-03		0.025	0.025	100.%	
CONTROL DUP ENVIRO WATER	LAB-ID #	Sample Concentr.	Spike Concentr.	QC Test Result	Pct (%) Recovery	RPD
Nitrogen, Total Persulfate-mg/L	0018543-04		1	0.97	97.%	2.%
Nitrogen, Total Persulfate-mg/L	0018565-04		1	0.93	93.%	3.2%
Phosphorus, Total-mg/L	0018547-04		2	2.04	102.%	1.%
Phosphorus, Total-mg/L	0018547-04		2	2.04	102.%	1.%
Phosphorus, Total low level-mg/L	0018531-04		0.025	0.025	100.%	0.%
Phosphorus, Total low level-mg/L	0018548-04		0.025	0.025	100.%	0.%
Phosphorus, Total low level-mg/L	0018548-04		0.025	0.025	100.%	0.%
Phosphorus, Total low level-mg/L	0018566-04		0.025	0.025	100.%	0.%
DUPLICATE ENVIRO WATER	LAB-ID #	Sample Concentr.	Spike Concentr.	QC Test Result	Pct (%) Recovery	RPD
Nitrogen, Total Persulfate-mg/L	2206823-10	0.64		0.64		0.%
Nitrogen, Total Persulfate-mg/L	2206823-30	0.8		0.8		0.%
Phosphorus, Total-mg/L	2207310-01	0.1		0.1		0.%
Phosphorus, Total-mg/L	2207310-01	0.1		0.1		0.%
Phosphorus, Total low level-mg/L	2206759-03	0.018		0.018		0.%
Phosphorus, Total low level-mg/L	2206823-15	0.027		0.027		0.%
Phosphorus, Total low level-mg/L	2206823-15	0.027		0.027		0.%
Phosphorus, Total low level-mg/L	2207522-02	0.005		0.005		0.%
Turbidity-NTU	2206823-34	0.6		0.59		1.7%

QUALITY CONTROL REPORT

Montana Environmental Laboratory LLC

Test Parameters

Order#: M2206823

MS	LAB-ID #	Sample Concentr.	Spike Concentr.	QC Test Result	Pct (%) Recovery	RPD
ENVIRO WATER						
Nitrogen, Total Persulfate-mg/L	2206823-10	0.64	1	1.58	94.%	
Nitrogen, Total Persulfate-mg/L	2206823-30	0.8	1	1.76	96.%	
Phosphorus, Total-mg/L	2207310-01	0.1	1	1.1	100.%	
Phosphorus, Total-mg/L	2207310-01	0.1	1	1.1	100.%	
Phosphorus, Total low level-mg/L	2206759-03	0.018	0.03	0.046	93.3%	
Phosphorus, Total low level-mg/L	2206823-15	0.027	0.03	0.056	96.7%	
Phosphorus, Total low level-mg/L	2206823-15	0.027	0.03	0.056	96.7%	
Phosphorus, Total low level-mg/L	2207522-02	0.005	0.03	0.035	100.%	
MSD	LAB-ID #	Sample Concentr.	Spike Concentr.	QC Test Result	Pct (%) Recovery	RPD
ENVIRO WATER						
Nitrogen, Total Persulfate-mg/L	2206823-10	0.64	1	1.58	94.%	0.0%
Nitrogen, Total Persulfate-mg/L	2206823-30	0.8	1	1.78	98.%	1.1%
Phosphorus, Total-mg/L	2207310-01	0.1	1	1.1	100.%	0.0%
Phosphorus, Total-mg/L	2207310-01	0.1	1	1.1	100.%	0.0%
Phosphorus, Total low level-mg/L	2206759-03	0.018	0.03	0.047	96.7%	2.2%
Phosphorus, Total low level-mg/L	2206823-15	0.027	0.03	0.056	96.7%	0.0%
Phosphorus, Total low level-mg/L	2206823-15	0.027	0.03	0.056	96.7%	0.0%
Phosphorus, Total low level-mg/L	2207522-02	0.005	0.03	0.035	100.%	0.0%
SRM	LAB-ID #	Sample Concentr.	Spike Concentr.	QC Test Result	Pct (%) Recovery	RPD
ENVIRO WATER						
Nitrogen, Total Persulfate-mg/L	0018543-05		3.24	3.12	96.3%	
Nitrogen, Total Persulfate-mg/L	0018565-05		3.235	2.94	90.9%	
Phosphorus, Total-mg/L	0018547-05		4.56	4.49	98.5%	
Phosphorus, Total-mg/L	0018547-05		4.56	4.49	98.5%	
Phosphorus, Total low level-mg/L	0018531-05		0.046	0.045	97.8%	
Phosphorus, Total low level-mg/L	0018548-05		0.046	0.049	106.5%	
Phosphorus, Total low level-mg/L	0018548-05		0.046	0.049	106.5%	
Phosphorus, Total low level-mg/L	0018566-05		0.046	0.046	100.%	